

I. Electrical – DC/Low Frequency

Parameter/Equipment	Range <sup>4</sup>	CMC <sup>2, 6</sup> ( $\pm$ )	Comments
DC Voltage <sup>3</sup> – Generate	(0 to 220) mV (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V  (220 to 1100) V  10.00 V 100.0 mV 1.0 V 10.0 V 100.0 V 1000.0 V	7 $\mu$ V/V + 0.5 $\mu$ V 4 $\mu$ V/V + 0.8 $\mu$ V 3 $\mu$ V/V + 5 $\mu$ V 3 $\mu$ V/V + 5 $\mu$ V 4 $\mu$ V/V + 50 $\mu$ V  6 $\mu$ V/V + 0.5 mV  0.5 $\mu$ V/V 0.9 $\mu$ V/V 0.8 $\mu$ V/V 0.75 $\mu$ V/V 0.8 $\mu$ V/V 0.9 $\mu$ V/V	Fluke 5720A       Fluke 732A array       Fluke 732A, 752A
DC Voltage <sup>3</sup> – Measure	(0 to 120) mV (0.1 to 1.2) V (1.0 to 12) V (10 to 120) V (100 to 1050) V  (0.5 to 10) kV (10 to 100) kV	5 $\mu$ V/V + 1 $\mu$ V 4 $\mu$ V/V + 1 $\mu$ V 4 $\mu$ V/V + 2 $\mu$ V 6 $\mu$ V/V + 30 $\mu$ V 18 $\mu$ V/V + 100 $\mu$ V  0.04 % IV + 0.03 V 0.075 % IV + 0.3 V	Agilent 3458A       Vitrek 4670A w/ HVL-100
DC Current <sup>3</sup> – Measure	(2 to 20) A  (10 to 100) A (30 to 300) A  (12 to 120) $\mu$ A (0.12 to 1.2) mA (1.2 to 12) mA (12 to 120) mA (0.12 to 1.2) A	0.01 % IV  0.05 % IV 0.1 % IV  20 $\mu$ A/A + 800 pA 20 $\mu$ A/A + 5.0 nA 20 $\mu$ A/A + 50 nA 35 $\mu$ A/A + 500 nA 0.011 % IV + 10 $\mu$ A	Fluke Y5020, HP 3458A  Guildline 9211A, Agilent 3458A  Agilent 3458A

Parameter/Equipment	Range <sup>4</sup>	CMC <sup>2, 6</sup> (±)	Comments
DC Current <sup>3</sup> – Generate	(0 to 220) µA (0.22 to 2.2) mA (2.2 to 22) mA (22 to 220) mA (0.22 to 2.2) A  (2.0 to 11) A	42 µA/A + 7 nA 35 µA/A + 8 nA 35 µA/A + 50 nA 45 µA/A + 0.8 µA 70 µA/A + 15 µA  0.034 % + 480 µA	Fluke 5720A  Fluke 5720A w/ Fluke 5725A
Resistance <sup>3</sup> – Measure	(0.0 to 12) Ω (10 to 120) Ω (0.10 to 1.2) kΩ (1.0 to 12) kΩ (10 to 120) kΩ (0.10 to 1.2) MΩ (1.0 to 12) MΩ (10 to 120) MΩ	15 µΩ/Ω + 50 µΩ 12 µΩ/Ω + 500 µΩ 10 µΩ/Ω + 500 µΩ 10 µΩ/Ω + 5 mΩ 10 µΩ/Ω + 50 mΩ 15 µΩ/Ω + 2 Ω 50 µΩ/Ω + 100 Ω 0.05 % IV + 1000 Ω	Agilent 3458A
Resistance <sup>3</sup> – Generate	10 000.00 Ω  1.0 Ω  1.0 Ω (Nominal) 1.9 Ω 10.0 Ω 19.0 Ω 100.0 Ω 190.0 Ω 1.0 kΩ 1.9 kΩ 10.0 kΩ 19.0 kΩ 100.0 kΩ 190.0 kΩ 1.0 MΩ 1.9 MΩ 10.0 MΩ 19.0 MΩ 100.0 MΩ	0.5 µΩ/Ω  0.5 µΩ/Ω  95 µΩ/Ω 95 µΩ/Ω 25 µΩ/Ω 25 µΩ/Ω 11 µΩ/Ω 11 µΩ/Ω 9 µΩ/Ω 9 µΩ/Ω 9 µΩ/Ω 9 µΩ/Ω 11 µΩ/Ω 11 µΩ/Ω 18 µΩ/Ω 19 µΩ/Ω 37 µΩ/Ω 47 µΩ/Ω 0.011 % IV	ESI SR-104  Leeds & Northrup 4210  Fluke 5720A

Parameter/Equipment	Range <sup>4</sup>	CMC <sup>2,5</sup> (±)	Comments
Resistance – Ratio Measurement	1:1 0.1:1 to 10:1 > 10:1 to 100:1	1.0 parts in $10^6$ 1.4 parts in $10^6$ 2.0 parts in $10^6$	MI 6010, MI 6000

Parameter/Range <sup>4</sup>	Frequency	CMC <sup>2,6</sup> (±)	Comments
AC Voltage <sup>3</sup> – Generate			
(0 to 2.2) mV	(10 to 19.99) Hz (20 to 39.99) Hz (0.04 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.027 % IV + 5 µV 0.011 % IV + 5 µV 90 µV/V + 5 µV 0.023 % IV + 5 µV 0.054 % IV + 6 µV 0.12 % IV + 12 µV 0.15 % IV + 25 µV 0.31 % IV + 25 µV	Fluke 5720A
(2.0 to 22) mV	(10 to 19.99) Hz (20 to 39.99) Hz (0.04 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.027 % IV + 5 µV 0.011 % IV + 5 µV 90 µV/V + 5 µV 0.023 % IV + 5 µV 0.054 % IV + 6 µV 0.12 % IV + 12 µV 0.15 % IV + 25 µV 0.31 % IV + 25 µV	
(20 to 220) mV	(10 to 19.99) Hz (20 to 39.99) Hz (0.04 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.027 % IV + 50 µV 0.01 % IV + 20 µV 50 µV/V + 10 µV 90 µV/V + 12 µV 0.013 % IV + 40 µV 0.042 % IV + 100 µV 0.11 % IV + 250 µV 0.18 % IV + 400 µV	
(0.20 to 2.2) V	(10 to 19.99) Hz (20 to 39.99) Hz (0.04 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.027 % IV + 500 µV 0.01 % IV + 200 µV 50 µV/V + 70 µV 90 µV/V + 120 µV 0.012 % IV + 250 µV 0.031 % IV + 800 µV 0.11 % IV + 2.5 mV 0.16 % IV + 4 mV	

Parameter/ Range <sup>4</sup>	Frequency	CMC <sup>2, 6</sup> ( $\pm$ )	Comments
AC Voltage <sup>3</sup> – Generate (cont)			
(2.0 to 22) V	(10 to 19.99) Hz (20 to 39.99) Hz (0.04 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.027 % IV + 5 mV 0.01 % IV + 2 mV 60 $\mu$ V/V + 7 mV 0.01 % IV + 1.2 mV 0.017 % IV + 3 mV 0.1 % IV + 20 mV 0.52 % IV + 50 mV 0.9 % IV + 100 mV	Fluke 5720A
(20 to 220) V	(10 to 19.99) Hz (20 to 39.99) Hz (0.04 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.027 % IV + 5 mV 0.01 % IV + 2 mV 60 $\mu$ V/V + 700 $\mu$ V 0.01 % IV + 12 mV 0.017 % IV + 3 mV	
(200 to 1100) V	(0.05 to 1) kHz	80 $\mu$ V/V + 4 mV	
(200 to 750) V	(30 to 50) kHz (50 to 100) kHz	0.036 % IV + 11 mV 0.13 % IV + 45 mV	Fluke 5720A w/ Fluke 5725A
(200 to 1100) V	(0.04 to 1) kHz (1 to 20) kHz (20 to 30) kHz	80 $\mu$ V/V + 4 mV 0.013 % IV + 6 mV 0.036 % IV + 11 mV	
AC Voltage <sup>3, 5</sup> – Measure			
(0 to 2.2) mV	(10 to 19.99) Hz (20 to 39.99) Hz (0.040 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.17 % IV + 1.3 $\mu$ V 0.074 % IV + 1.3 $\mu$ V 0.042 % IV + 1.3 $\mu$ V 0.081 % IV + 2.0 $\mu$ V 0.12 % IV + 2.5 $\mu$ V 0.23 % IV + 4.0 $\mu$ V 0.24 % IV + 8.0 $\mu$ V 0.35 % IV + 8.0 $\mu$ V	Fluke 5790A
(2.2 to 7.0) mV	(10 to 19.99) Hz (20 to 39.99) Hz (0.040 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.085 % IV + 1.3 $\mu$ V 0.037 % IV + 1.3 $\mu$ V 0.021 % IV + 1.3 $\mu$ V 0.04 % IV + 2.0 $\mu$ V 0.06 % IV + 2.5 $\mu$ V 0.12 % IV + 4.0 $\mu$ V 0.13 % IV + 8.0 $\mu$ V 0.23 % IV + 8.0 $\mu$ V	

Parameter/ Range <sup>4</sup>	Frequency	CMC <sup>2, 6</sup> ( $\pm$ )	Comments
AC Voltage <sup>3, 5</sup> – Measure (cont)			
(7.0 to 22.0) mV	(10 to 19.99) Hz (20 to 39.99) Hz (0.040 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.029 % IV + 1.3 $\mu$ V 0.019 % IV + 1.3 $\mu$ V 0.011 % IV + 1.3 $\mu$ V 0.021 % IV + 2.0 $\mu$ V 0.031 % IV + 2.5 $\mu$ V 0.081 % IV + 4.0 $\mu$ V 0.089 % IV + 8.0 $\mu$ V 0.17 % IV + 8.0 $\mu$ V	Fluke 5790A
(22 to 70) mV	(10 to 19.99) Hz (20 to 39.99) Hz (0.040 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.024 % IV + 1.5 $\mu$ V 0.012 % IV + 1.5 $\mu$ V 65 $\mu$ V/V + 1.5 $\mu$ V 0.013 % IV + 2.0 $\mu$ V 0.026 % IV + 2.5 $\mu$ V 0.051 % IV + 4.0 $\mu$ V 0.067 % IV + 8.0 $\mu$ V 0.11 % IV + 8.0 $\mu$ V	
(70 to 220) mV	(10 to 19.99) Hz (20 to 39.99) Hz (0.040 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.021 % IV + 1.5 $\mu$ V 85 $\mu$ V/V + 1.5 $\mu$ V 38 $\mu$ V/V + 1.5 $\mu$ V 69 $\mu$ V/V + 2.0 $\mu$ V 0.016 % IV + 2.5 $\mu$ V 0.025 % IV + 4.0 $\mu$ V 0.038 % IV + 8.0 $\mu$ V 0.1 % IV + 8.0 $\mu$ V	
(220 to 700) mV	(10 to 19.99) Hz (20 to 39.99) Hz (0.040 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.021 % IV + 1.5 $\mu$ V 76 $\mu$ V/V + 1.5 $\mu$ V 33 $\mu$ V/V + 1.5 $\mu$ V 51 $\mu$ V/V + 2.0 $\mu$ V 79 $\mu$ V/V + 2.5 $\mu$ V 0.018 % IV + 4.0 $\mu$ V 0.03 % IV + 8.0 $\mu$ V 0.096 % IV + 8.0 $\mu$ V	
(0.7 to 2.2) V	(10 to 19.99) Hz (20 to 39.99) Hz (0.040 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.02 % IV 66 $\mu$ V/V 24 $\mu$ V/V 46 $\mu$ V/V 71 $\mu$ V/V 0.016 % IV 0.026 % IV 0.09 % IV	

Parameter/ Range <sup>4</sup>	Frequency	CMC <sup>2, 6</sup> ( $\pm$ )	Comments
AC Voltage <sup>3, 5</sup> – Measure (cont)			
(2.2 to 7.0) V	(10 to 19.99) Hz (20 to 39.99) Hz (0.040 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.02 % IV 67 $\mu$ V/V 24 $\mu$ V/V 48 $\mu$ V/V 81 $\mu$ V/V 0.019 % IV 0.04 % IV 0.12 % IV	Fluke 5790A
(70 to 220) V	(10 to 19.99) Hz (20 to 39.99) Hz (0.04 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz	0.02 % IV 68 $\mu$ V/V 31 $\mu$ V/V 69 $\mu$ V/V 98 $\mu$ V/V 0.021 % IV 0.05 % IV	
(220 to 700) V	(10 to 19.99) Hz (20 to 39.99) Hz (0.04 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.02 % IV 96 $\mu$ V/V 39 $\mu$ V/V 0.012 % IV 0.04 % IV	
(700 to 1000) V	(10 to 19.99) Hz (20 to 39.99) Hz (0.040 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.02 % IV 96 $\mu$ V/V 37 $\mu$ V/V 0.012 % IV 0.04 % IV	
(0 to 10) kV	60 Hz	0.15 % IV + 0.1 V	Vitrek 4700A w/ HVL-100
(10 to 75) kV	60 Hz	0.15 % IV + 0.6 V	

Parameter/ Range	Frequency	CMC <sup>2, 6</sup> ( $\pm$ )	Comments
AC Current <sup>3</sup> – Measure			
(0 to 120) $\mu$ A	(10 to 19.99) Hz (20 to 44.99) Hz (45 to 100) Hz (0.1 to 1) kHz	0.4 % IV + 30 pA 0.15 % IV + 30 pA 0.06 % IV + 30 pA 0.06 % IV + 30 pA	Agilent 3458A
(0.12 to 1.2) mA	(10 to 19.99) Hz (20 to 44.99) Hz (45 to 100) Hz (0.1 to 5) kHz	0.4 % IV + 200 pA 0.15 % IV + 200 pA 0.06 % IV + 200 pA 0.03 % IV + 200 pA	
(1.2 to 12) mA	(10 to 19.99) Hz (20 to 44.99) Hz (45 to 100) Hz (0.1 to 5) kHz	0.4 % IV + 2 $\mu$ A 0.15 % IV + 2 $\mu$ A 0.06 % IV + 2 $\mu$ A 0.03 % IV + 2 $\mu$ A	
(12 to 120) mA	(10 to 19.99) Hz (20 to 44.99) Hz (45 to 100) Hz (0.1 to 5) kHz	0.15 % IV + 20 $\mu$ A 0.06 % IV + 20 $\mu$ A 0.03 % IV + 20 $\mu$ A 0.06 % IV + 20 $\mu$ A	
(0.12 to 2.2) A	(10 to 19.99) Hz (20 to 44.99) Hz (45 to 100) Hz (0.1 to 5) kHz	0.4 % IV + 200 $\mu$ A 0.16 % IV + 200 $\mu$ A 0.08 % IV + 200 $\mu$ A 0.1 % IV + 200 $\mu$ A	
(2 to 20) A	(50 to 100) Hz 300 Hz 1 kHz 3 kHz 4 kHz 5 kHz	0.016 % IV 0.019 % IV 0.027 % IV 0.051 % IV 0.063 % IV 0.075 % IV	Fluke Y5020, HP 3458A
(20 to 1200) A	60 Hz 400 Hz	0.05 % IV 0.09 % IV	Weston 327 current transformer, Fluke Y5020, HP 3458A

Parameter/ Range	Frequency	CMC <sup>2, 6</sup> (±)	Comments
AC Current <sup>3</sup> – Generate			
(0 to 220) µA	(10 to 19.99) Hz (20 to 39.99) Hz (0.04 to 1) kHz (1 to 5) kHz (5 to 10) kHz	0.036 % IV + 2.1 µA 0.023 % IV + 2.1 µA 0.017 % IV + 2.1 µA 0.042 % IV + 2.1 µA 0.15 % IV + 2.1 µA	Fluke 5720A
(0.2 to 2.2) mA	(10 to 19.99) Hz (20 to 39.99) Hz (0.04 to 1) kHz (1 to 5) kHz (5 to 10) kHz	0.036 % IV + 2.1 µA 0.023 % IV + 2.1 µA 0.017 % IV + 2.1 µA 0.029 % IV + 2.2 µA 0.15 % IV + 3.1 µA	
(2.0 to 22) mA	(10 to 19.99) Hz (20 to 39.99) Hz (0.04 to 1) kHz (1 to 5) kHz (5 to 10) kHz	0.028 % IV + 0.5 µA 0.018 % IV + 0.4 µA 0.013 % IV + 0.4 µA 0.022 % IV + 0.7 µA 0.11 % IV + 6 µA	
(20 to 220) mA	(10 to 19.99) Hz (20 to 39.99) Hz (0.04 to 1) kHz (1 to 5) kHz (5 to 10) kHz	0.028 % IV + 5 µA 0.018 % IV + 4 µA 0.013 % IV + 3 µA 0.022 % IV + 4 µA 0.11 % IV + 12 µA	
(0.20 to 2.2) A	(0.02 to 1) kHz (1 to 5) kHz (5 to 10) kHz	0.03 % IV + 40 µA 0.046 % IV + 100 µA 0.7 % IV + 200 µA	
(2.0 to 11) A	(0.04 to 1) kHz (1 to 5) kHz (5 to 10) kHz	0.04 % IV + 170 µA 0.085 % IV + 380 µA 0.33 % + 750 µA	Fluke 5720A w/ Fluke 5725A
(2 to 20) A	50 to 100 Hz 300 Hz 1 kHz 3 kHz 4 kHz 5 kHz	0.016 % IV 0.019 % IV 0.027 % IV 0.051 % IV 0.063 % IV 0.075 % IV	Fluke Y5020, HP 3458A, Fluke 5725A
(20 to 100) A	60 Hz	0.05 % IV	Fluke Y5020, HP 3458A, Weston 327 CT, Vahalla 2555A

Parameter/ Range <sup>4</sup>	Frequency	CMC <sup>2, 6</sup> ( $\pm$ )	Comments
Capacitance <sup>3</sup> – Generate			
1000 pF	1000 Hz	5.0 $\mu$ F/F	GenRad 1404A Fluke
(330 to 500) pF	(50 to 1000) Hz	0.5 % IV + 10 pF	
(0.5 to 1.1) nF		0.5 % IV + 10 pF	5500A
(1.1 to 3.3) nF		0.5 % IV + 10 pF	
(3.3 to 11) nF		0.5 % IV + 10 pF	
(11 to 33) nF		0.25 % IV + 100 pF	
(33 to 110) nF		0.25 % IV + 100 pF	
(110 to 330) nF		0.25 % IV + 300 pF	
(0.330 to 1.10) $\mu$ F		0.25 % IV + 1 nF	
(1.1 to 3.3) $\mu$ F		0.35 % IV + 3 nF	
(3.3 to 11) $\mu$ F	(50 to 400) Hz	0.35 % IV + 10 nF	
(11 to 33) $\mu$ F		0.4 % IV + 30 nF	
(33 to 110) $\mu$ F	(50 to 200) Hz	0.5 % IV + 100 nF	
(110 to 330) $\mu$ F	(50 to 100) Hz	0.7 % IV + 300 nF 1 % IV + 300 nF	
(0.330 to 1.1) mF			
Capacitance <sup>3</sup> – Measure			
(0.1 to 1000) pF	1 kHz	0.01 % IV + 30 fF	GenRad 1620A Fluke
(0 to 1) nF	(0.050 to 2) kHz	0.1 % IV + 10 fF	
(1 to 10) nF		0.1 % IV + 100 fF	PM6304C
(10 to 100) nF		0.1 % IV + 1 pF	
(0.1 to 1) $\mu$ F		0.1 % IV + 10 pF	
(1 to 10) $\mu$ F		0.1 % IV + 100 pF	
(10 to 100) $\mu$ F		0.1 % IV + 1 nF	
(0.1 to 1) mF		0.1 % IV + 10 nF	
(1 to 10) mF		0.1 % IV + 100 nF	
(10 to 100) mF		0.1 % IV + 1 $\mu$ F	
(0.1 to 1) F		0.1 % IV + 10 $\mu$ F	
(1 to 10) F		0.1 % IV + 100 $\mu$ F	
(10 to 31.8) F		0.1 % IV + 1 mF	

Parameter/ Range	Frequency	CMC <sup>2, 6</sup> (±)	Comments
Inductance <sup>3</sup> – Measure			
(0.00 to 1.00) mH (1.00 to 10) mH (10 to 100) mH (0.100 to 1.00) mH (1 to 9.9999) H (10 to 99.999) H (100 to 999.99) H	(0.050 to 2) kHz	0.1 % IV + 10 nH 0.1 % IV + 100 nH 0.1 % IV + 1 µH 0.1 % IV + 10 µH 0.1 % IV + 100 µH 0.1 % IV + 1 mH 0.1 % IV + 10 mH	Fluke PM6304C
Oscilloscopes <sup>3</sup> – Amplitude			
DC – DC Signal 50 Ω Load 1 MΩ Load	(0 to ± 2.2) V (0 to ± 33) V	0.25 % IV + 100 µV 0.25 % IV + 100 µV	Fluke 5500A w/ SC600 scope option
Amplitude – Square Wave (Peak to Peak) – 50 Ω Load	± 1.8 mV to ± 2.2 V ± 1.8 mV to ± 55 V	0.25 % IV + 100 µV 0.25 % IV + 100 µV	
Amplitude – Square Wave 1 MΩ Load – (10 to 100) Hz (0.1 to 10) kHz	(95 to 105) V	1.5 % IV ± 100 µV 0.5 % IV ± 100 µV	
Leveled Sine Wave – (Into 50 Ω Load)	50 kHz reference 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz	2 % IV + 300 µV 3.5 % IV + 300 µV 4 % IV + 300 µV 6 % IV + 300 µV	
Flatness @ 50 kHz Reference	50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz	1.5 % IV + 100 µV 2 % IV + 100 µV 4 % IV + 100 µV	
Time Marker – (Into 50 Ω Load)	5 s to 50 ms 20 ms to 1 ns	(25 + 1000t) µs/s 2.5 µs/s	t is the numerical value of the time in seconds
Edge Spec (Rise Time)	≤ 1000 ps	(+0ps / -900ps)	Referenced to 50 kHz
Amplitude Flatness	0.1 Hz to 50 kHz (0.050 to 100) MHz (100 to 250) MHz (250 to 550) MHz	1.5 % IV 1.5 % IV 3 % IV 4 % IV	

Parameter/ Range	Range	CMC <sup>2</sup> (±)	Comments
Thermocouple Simulation – Generate and Measure			
Type E	(-250 to -100) °C (-100 to -25) °C (-25 to 350) °C (350 to 650) °C (650 to 1000) °C	0.5 °C 0.16 °C 0.14 °C 0.16 °C 0.21 °C	Fluke 5500A
Type J	(-210 to -100) °C (-100 to -30) °C (-30 to 150) °C (150 to 760) °C (760 to 1200) °C	0.27 °C 0.16 °C 0.14 °C 0.17 °C 0.23 °C	
Type K	(-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 1000) °C (1000 to 1372) °C	0.33 °C 0.18 °C 0.16 °C 0.26 °C 0.4 °C	
Type R	(0 to 250) °C (250 to 400) °C (400 to 1000) °C (1000 to 1767) °C	0.57 °C 0.35 °C 0.33 °C 0.4 °C	
Type S	(0 to 250 °C (250 to 1000 °C (1000 to 1400 °C (1400 to 1767 °C	0.47 °C 0.36 °C 0.37 °C 0.46 °C	
Type T	(-250 to -150) °C (-150 to 0) °C (0 to 120) °C (120 to 400) °C	0.63 °C 0.24 °C 0.16 °C 0.14 °C	